





## Kitemark™ Certificate

This is to certify that: Albert Browne Ltd

Chancery House Rayns Way

Watermead Business Park

Syston, Leicester

LE7 1PF

United Kingdom

Holds Certificate Number: KM 60358

In respect of:

ISO 11140-1 Sterilization of Health care products. Chemical indicators

This issues the right and licence to use the Kitemark in accordance with the Kitemark Terms and Conditions governing the use of the Kitemark, as may be updated from time to time by BSI Assurance UK Ltd (the "Conditions"). All defined terms in this Certificate shall have the same meaning as in the Conditions.

The use of the Kitemark is authorized in respect of the Product(s) detailed on this Certificate provided at or from the above address.

For and on behalf of BSI: Shahm Barhom, Group Product Certification Director

First Issued: 2001-10-09 Effective Date: 2025-02-24 Latest Issue: 2025-02-24 Expiry Date: 2025-12-31

Page: 1 of 2

## Kitemark™ Certificate

No. KM 60358

**Product Specification** 

BS EN ISO 11140-1:2014

Sterilization of health care products. Chemical indicators. General requirements.

## TST Control™ Integrators – Type 6 Emulating Indicators:

Ink Code	Ink Timing (Stated Value in Decimal Minutes)			Product Code
	134°C	121 °C	270 °F	
2396a	3.5 min	N/A	N/A	2317, 2340, 3727 (self-adhesive), 3103, V3502060, V3504065
2396c	4.0 min	12.0 min	N/A	2341, 3104, V3502160, 2375, 2341S, 2341GCC
2396b	5.3 min	15.0 min	N/A	2302, 2342, 3726 (self-adhesive), 3105, V3502065, V3504068
2396g	7.0 min	20.0 min	N/A	3725 (self-adhesive), 3760, 3107, V3502070
2396L	9.0 min	25.0 min	N/A	3702
3706-01	18.0 min	N/A	N/A	3706
2396d	N/A	N/A	3.0 min	Currently none
2396L	N/A	N/A	10.0 min	Currently none
7320	18.0 min	N/A	N/A	3108 (self-adhesive)

First Issued: 2001-10-09 Effective Date: 2025-02-24 Latest Issue: 2025-02-24 Expiry Date: 2025-12-31

Page: 2 of 2

This certificate has been issued by and remains the property of BSI Assurance UK Ltd, Kitemark Court, Davy Avenue, Knowlhill, Milton Keynes MK5 8PP, United Kingdom and should be returned immediately upon request.

To check its validity telephone +44 (0) 345 080 9000. An electronic certificate can be authenticated online.