

SGS Portugal, S.A.
Industrial Services
Pólo Tecnológico de Lisboa,
Rua Cesina Adães Bermudes,
Lote 11, nº1 1600-604 Lisboa – Portugal
T. (351) 21 710 42 00
F. (351) 21 715 75 26



Certificate of Conformity

EU Type Examination (Production Type)

According to Directive 2014/68/EU

Certificate N°: **PTC25.03646.5308**

Date: **2025.06.03**

Name and Address
of the Manufacturer: **PROHS - Equipamento Hospitalar e
Serviços Associados, S.A.
Rua do Castanhal N°316, Z.Industrial Maia I,
Sector II, 4475-122 Maia
Portugal**

SGS Portugal SA as a notified body for the scope of Pressure Equipment Directive 2014/68/EU, certifies that the Pressure Equipment design, production, inspection and assembly, is in compliance with the requirements of the Pressure Equipment Directive stated, according to module B (Production Type).

NOBO 1155


Assessed According to
Directive 2014/68/EU: **Module B (Production)
(Assembly)**

Inspection Report N°: **PTR25.07396.5308 & PTR25.07397.5308**

Equipment ID: **Sistema de Esterilização ;
Vol: 110L / 39L / 30L (Chamber / Jacket / GV) ;
SN: 1951 / 9696 (EH / GV) ;
Model: FJ110 1PD & GV 320-30;
Variants: FJ110 2PD & GV320-30; FJ100
1/2PD & GV320-30; FJ70 1/2PD & GV320-30**

PED Data: **PS [bar]: -1,0 / 3,0 (Chamber) ; ATM / 3.0 (Jacket) ;
5,0 (GV)
TS [°C]: -10 / 140 (Chamber & Jacket) ; -10 / 165 (GV)
Category: III
Fluid/Group: Gas – Gr.2 (EH & GV)
PT [bar]: 3,0 / 5,0 (Abertura de válvulas EH/GV)
Date: 2025-06-18
Year of Manufacture: 2025**

Valid Until: **2035.07.02**



Carlos Felgueiras
Technical Responsible

