

EU Type Examination (Production Type)

According to Directive 2014/68/EU

Certificate N°: **PTC24.03856.5308**

Date: **2024.06.26**

Name and Address
of the Manufacturer:

**PROHS - Equipamento Hospitalar e
Serviços Associados, S.A.
Rua do Castanhal N°316, Z.Industrial da
Maia I, Sector II, 4475-122, Maia
Portugal**



C0039
ISO/IEC 17065
Produtos

Certificate of Conformity

SGS Portugal SA as a notified body for the scope of Pressure Equipment Directive 2014/68/EU, certifies that the Pressure Equipment design, construction and inspection, is in compliance with the requirements of the Pressure Equipment Directive stated, according module B (Production Type).



1155

Assessed According to
Directive 2014/68/EU:

Module B (Production)

Inspection Report N°: **PTR24.07619.5308 & PTR24.07620.5308**

Equipment ID:

**Horizontal Sterilizer 930L ; SN: 1909 ;
Vol: 930L / 110L (Chamber / Jacket) ;
Model: FJ 930 1PD ;
Variants: FJ 930 2PD; FJ 780 1/2PD; FJ 740
1/2PD; FJ 640 1/2PD; FJ 610 1/2PD; FJ 490
1/2PD; FJ 340 1/2PD**

PED Data:

**PS [bar]: -1.0 / 3,0 (Chamber) ; Atm / 3,0 (Jacket)
TS [°C]: -10 / 140 (Chamber & Jacket)
Category: III
Fluid/Group: Gas - Gr.2 (Chamber & Jacket)
PT [bar]: 7,0 / 9,3 (Chamber / Jacket)
Date: 2024-05-28
Year of Manufacture: 2024**

Valid Until: **2034.06.25**



**José Fabião
Technical Manager**

