

**SGS Portugal, S.A.**  
Industrial Services  
Pólo Tecnológico de Lisboa,  
Rua Cesina Adães Bermudes,  
Lote 11, nº1 1600-604 Lisboa – Portugal  
T. (351) 21 710 42 00  
F. (351) 21 715 75 26



C0039  
ISO/IEC 17065  
Produtos

## EU Type Examination (Production Type)

According to Directive 2014/68/EU

Certificate Nº: **PTC25.03647.5308**

Date: **2025.06.03**

Name and Address  
of the Manufacturer: **PROHS - Equipamento Hospitalar e  
Serviços Associados, S.A.  
Rua do Castanhal Nº316, Z.Industrial Maia I,  
Sector II, 4475-122 Maia  
Portugal**

SGS Portugal SA as a notified body for the scope of Pressure Equipment Directive 2014/68/EU, certifies that the Pressure Equipment design, production, and inspection, is in compliance with the requirements of the Pressure Equipment Directive stated, according to module B (Production Type).

Certificate of Conformity

**NOBO 1155**

Assessed According to  
Directive 2014/68/EU:

**Module B (Production)**

Inspection Report Nº: **PTR25.07398.5308 & PTR25.07399.5308**

Equipment ID: **Esterilizador Horizontal 360L ; SN: 24135 ;  
Vol: 360L / 54L (Chamber / Jacket) ;  
Model: FJ360 1PD; Variants: FJ360 2PD;  
FJ250 1/2PD; FJ175 1/2PD**

PED Data: **PS [bar]: -1,0 / 3,0 (Chamber) ; ATM / 3.0 (Jacket)  
TS [°C]: -10 / 140 (Chamber & Jacket)  
Category: III  
Fluid/Group: Gas – Gr.2 (Chamber & Jacket)  
PT [bar]: 7,0 / 9,3 (Chamber / Jacket)  
Date: 2025-02-24  
Year of Manufacture: 2025**

Valid Until: **2035.06.02**



Carlos Felgueiras  
Technical Responsible

