

SGS Portugal, S.A.
Industrial Services
Pólo Tecnológico de Lisboa,
Rua Cesina Adães Bermudes,
Lote 11, nº1 1600-604 Lisboa – Portugal
T. (351) 21 710 42 00
F. (351) 21 715 75 26



EU Type Examination (Production Type)

According to Directive 2014/68/EU

Certificate Nº: **PTC25.03645.5308**

Date: **2025.06.03**

Name and Address
of the Manufacturer: **PROHS - Equipamento Hospitalar e
Serviços Associados, S.A.
Rua do Castanhal Nº316, Z.Industrial Maia I,
Sector II, 4475-122 Maia
Portugal**

SGS Portugal SA as a notified body for the scope of Pressure Equipment Directive 2014/68/EU, certifies that the Pressure Equipment design, production, and inspection, is in compliance with the requirements of the Pressure Equipment Directive stated, according to module B (Production Type).

Certificate of Conformity

NOBO 1155

Assessed According to
Directive 2014/68/EU:

Module B (Production)

Inspection Report Nº: **PTR25.07394.5308 & PTR25.07395.5308**

Equipment ID: **Esterilizador Horizontal 110L ; SN: 1951 ;
Vol: 110L / 39L (Chamber / Jacket) ;
Model: FJ110 1PD ; Variants: FJ110 2PD;
FJ100 1/2PD; FJ70 1/2PD**

**PS [bar]: -1,0 / 3,0 (Chamber) ; ATM / 3.0 (Jacket)
TS [°C]: -10 / 140 (Chamber & Jacket)
Category: III**

PED Data: **Fluid/Group: Gas – Gr.2 (Chamber & Jacket)
PT [bar]: 7,0 / 9,3 (Chamber / Jacket)
Date: 2025-03-28
Year of Manufacture: 2025**

Valid until: **2035.06.02**



Carlos Felgueiras
Technical Responsible

